

Supplement

Title: 2AC Inst. Sht. Supplement Issue: 2
Print Date: August 2010 Issue Date: 1/11
Part Number: Page Count: 1
Revision/Date:

This supplement contains information necessary to ensure the accuracy of the above manual.

Change #1, 56708

Replace **Operating** Specification, with the following:

Operating: North America: 90 V ac to 1000 V ac
Europe: 200 V ac to 1000 V ac

Change #2

For the Spanish Instructions, under Advertencia:

Change:

- **No lo utilice si el probador parece dañado o si no funciona correctamente. En caso de duda, solicite servicio técnico de mantenimiento para el comprobador.**

To:

- **No lo utilice si el probador parece dañado o si no funciona correctamente. Examine específicamente la punta de la sonda para comprobar si tiene grietas o roturas antes de usarla. En caso de duda, lleve el probador a reparar.**
- **No aplicar más que la tensión nominal indicada en el comprobador.**